

Annexure -1 Lab Equipment

S.No.	Item Description	Uom	Detail specification	Prefferable Make	Application
1	MOISTURE ANALYSER	NOS	KF Moisture Content Analysis package (P/N-AP0014.KF1121) Range 1 To 10000 ppm Moisture Content. Diaphragm type. With complete Consumable and accessories required for one year operations.	Metrohm/Mettler toledo	Oil Testing