

**Corrigendum-03 for Tender No- ITPS/unit 3&4/ 2018-19/099 dtd 25/10/2018
(Supply of CEQMS)**

Sl. No.	Clause No. / Page No.	Existing	Amended
1	Page no 5: Special Terms and Condition, Point no 3.	<ul style="list-style-type: none"> ➤ Only insitu type analyser will be accepted. 	<ul style="list-style-type: none"> ➤ Insitu type analyser will be accepted. However, in case of extractive type analyser only TOC type extractive analyser with NDIR technology will be accepted.
2	Annexure 1	Technical Specification	Revised Specification as attached
3	NIT	Bid Submission Date : 27.12.2018	Bid Submission Date : 01.02.2019
All other terms & condition of RFP#ITPS/Unit 3&4/2018-19/099 dtd 25/10/2018) will remain unchanged.			